An official EU website



- ♠ Safety Gate home (https://ec.europa.eu/consumers/consumers_safety/safety_products/rapex/alerts/repository/content/pages/rapex/index_en.htm)
- I Back to report listings (https://ec.europa.eu/consumers/consumers safety/safety products/rapex/alerts/?event=main.listNotifications&year=2020&lng=er
- Most recent alerts (https://ec.europa.eu/consumers/consumers_safety/safety_products/rapex/alerts/?event=main.immediatlyPublishedNotifications)
- Q Search alerts (https://ec.europa.eu/consumers/consumers_safety/safety_products/rapex/alerts/?event=main.search&Ing=en)
- IPSW 2018 (https://ec.europa.eu/consumers/consumers_safety/safety_products/rapex/alerts/?event=ipsw:home)

Safety Gate: Rapid Alert System for dangerous non-food products - most recent alerts

Alert number: A12/00776/20 Product: Particle filter mask Name: Cubrebocas anatomico Risk type: Health risk / other Category: Protective equipment

Brand: Protekcia

Type / number of model: CVE: CUP02, Hindernis.mx, Registry number: 1673C2017 SSA

Type of alert: Products with serious risks

The particle/filter retention of the material is insufficient (measured value ≤ 64 %) and the mask does not properly adapt to the face. Consequently, an excessive amount of particles or microorganisms might pass through the mask, increasing the risk of infection if not combined with additional protective measures. / The product does not comply with the Personal Protective Equipment Regulation.

Measures ordered by public authorities (to: Other): Ban on the marketing of the product and any accompanying measures

Description: Anatomic disposable single use mouth protective mask.

Packaging description: The product is sold in a transparent plastic bag containing 100 units and a paper label in black and blue printing.

Barcode: CUP02 Counterfeit: NO

Batch number: Expiry date: 1 February 2025, H101CA0920

Country of origin: Mexico
Alert submitted by: Belgium







