

Alert number

[A12/00642/21](#)

Product

Particle filter mask



Risk type

Health risk / other

The particle/filter retention of the material is insufficient (measured values: as low as 5%). The total filtration capacity of the mask is insufficient (measured values: as low as 45%). Consequently, an excessive amount of particles or microorganisms might pass through the mask; thus, even if combined with other recommended measures, it may not properly protect.

The product does not comply with the PPE Regulation and the relevant European standard EN 149.

Category

Protective equipment

Type

Consumer

Description

Black particle filter mask with small valve.

Packaging description

The product is sold packed in a black and red plastic wrapping, containing a single item.

Brand

Dr. Frei

Name

Protection Mask

Batch number

exp. 2022

Barcode

3800732484840

Weekly report number

Report-2021-19

Alert submitted by	Iceland
Is the product counterfeit?	Unknown
Country of origin	Unknown
Measures taken by economic operators	Recall of the product from end users Importer
Date of entry into force	01/04/2021