

## Safety Gate Alerts

Alert number

A12/00779/21

Product

Particle filter mask



Risk type **Health risk / other** 

The particle/filter retention of the material is insufficient (measured values: as low as 73%) and the total filtration capacity of the mask is insufficient (measured mean values: as low as 60%). Furthermore, the mask does not fit correctly to the face of the user. Consequently, an excessive amount of particles or microorganisms might pass through the mask; thus, even if combined with other recommended measures, it may not properly protect.

The product does not comply with the Personal Protective Equipment Regulation and with the relevant European standard EN 149.

Category

**Protective equipment** 

Type

Consumer

Description

White respiratory protective half-face mask.

Packaging description

The product is sold in a white carboard box containing 10 pieces.

Brand

Unknown

Name

KN95 mask

Weekly report number

Report-2021-22

Alert submitted by

**Iceland** 

Is the product counterfeit?

Unknown

09/06/2021 Page 1

Country of origin **Unknown** 

Measures taken by economic Recall of the product from end users

operators **Distributor** 

Date of entry into force 23/03/2021

Company recall code - <a href="https://rv.is/thjonusta-/innkollun-voru/">https://rv.is/thjonusta-/innkollun-voru/</a>

09/06/2021 Page 2