

Alert number

A12/00962/21

Product

Particle filter mask



Risk type

Health risk / other

The particle/filter retention of the material is insufficient (measured values: as low as 72.5%). Consequently, the product does not fulfil the health and safety requirements; thus, even if combined with other recommended measures, it may not properly protect.

The product does not comply with the Personal Protective Equipment (PPE) Regulation and with the relevant European standard EN 149.

Category

Protective equipment

Type

Professional

Description

Protective respiratory filtration half-face mask. The product was sold online, in particular via Amazon.

Packaging description

The product is sold in transparent plastic pouches.

Brand

DECADE

Name

Respirator FFP2 respiratory protective mask FFP2

Type / number of model

univent-medical

Batch number

Art-Nr. 1488

Barcode

4250397303499

Weekly report number

Report-2021-26

| | |
|--|--|
| Alert submitted by | Germany |
| Is the product counterfeit? | No |
| Country of origin | Germany |
| Measures ordered by public authorities | Ban on the marketing of the product and any accompanying measures Other |
| Date of entry into force | 21/05/2021 |