

Alert number

**[A12/01064/21](#)**

Product

**Particle filter mask**



Risk type

**Health risk / other**

**The total filtration capacity of the mask and the particle/filter retention of the material are insufficient (measured mean values as low as 78%, and measured filter penetration values as low as 89%, respectively). Consequently, the product does not fulfil the health and safety requirements; thus, even if combined with other recommended measures, it may not properly protect.**

**The product does not comply with the Personal Protective Equipment (PPE) Regulation and with the relevant European standard EN 149.**

Category

**Protective equipment**

Type

**Consumer**

Description

**White respiratory face mask with ear loops. The product was sold online, in particular via Amazon.**

Packaging description

**The product is sold individually packaged or in larger packaging units. The claim FFP2 and the CE marking are printed both on the box and on the individual product.**

Brand

**Univent medical**

Name

**atemicus pro**

Type / number of model

**2001**

Barcode

**4260688380047**

Weekly report number **Report-2021-29**

Alert submitted by **Germany**

Is the product counterfeit? **No**

Country of origin **Germany**

Measures ordered by public authorities **Recall of the product from end users  
Other**

Measures taken by economic operators **Stop of sales  
Manufacturer**